

Specifications of Cholelithoscope for the Department of General Surgery.

Optical System	Field of View	120°
	Direction of View	Forward viewing
	Depth of Field	3 -100 mm
Distal End	Outer Diameter	4.8 mm
Bending Section	Angulation Range	Up 160° / Down 130°
Insertion Tube	Outer Diameter	5.2 mm
Length	Working Length	380 mm
	Total Length	660 mm
Biopsy Forceps	Minimum Visible Distance	3 mm from distal end
Instrument Channel	Inner Diameter	2.0 mm
Sterilization	Compatible Methods	V-Pro/EtQ/Sterrad
Monitor	Compatible	Compatible

Dr. Ajay Kumar
Professor & Head
Department of Surgery
HIMSR New Delhi
DMC - 24837