


Specifications for Laparoscopic CO₂ Insufflator

High Flow CO₂ Gas Insufflator unit

- Should be digital, microprocessor controlled & automatic type insufflator of at least 45l.
- Should be suitable for all kinds of laparoscopic surgeries and should be capable to handle the neonatal, pediatric, adult and bariatric surgeries.
- Should have large display on front panel for information about set pressure, current pressure and CO₂ flow rate.
- It should give both visible and acoustic alarms for abnormal pressure, tube kinking and gas cylinder status.
- Powerful Insufflation flow rate of 45L or more. ? per/?
- Need to provide Pin and DIN type CO₂ hose pipe which can be connected to respective type CO₂ Cylinders.
- Should be supplied with all required accessories including the insufflation tubing's, wrench etc.
- Company should have satisfactory presence in India since last 10 years for Sales and Service.
- Should be US-FDA approved.


Dr. Ajay Triakral
Professor & Head
Department of Surgery
HIMSR New Delhi
DMC - 24837